



Product Service

# Attestation of Conformity

No. E8A 115826 0007 Rev. 00

**Holder of Certificate:** **sensor system technology co., ltd**  
Rm5009, Rm5010, Rm5011, 5F, Rm4009, 4F  
Ansan digitalpark, 81, Neungan-ro  
Danwon-gu, Ansan-si, Gyeonggi-do 15425  
REPUBLIC OF KOREA

**Name of Object:** **Pressure Meters**  
**(Pressure Transmitter)**

**Model(s):** **PUB**

**Description of Object:** Rated input voltage: 12 - 28 V d.c.  
Rated output: 4 - 20 mA

**Tested according to:** EN 61326-1:2013  
EN IEC 61326-1:2021  
EN 55011:2009/A1:2010  
EN 55011:2016/A11:2020

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** KR22-YEC0237

**Date,** 2022-11-17

( Laurent Yuan )