



Product Service

Attestation of Conformity

No. E8A 115826 0012 Rev. 00

Holder of Attestation: **sensor system technology co., ltd**
Rm5009, Rm5010, Rm5011, 5F, Rm4009, 4F
Ansan digitalpark, 81, Neungan-ro
Danwon-gu, Ansan-si, Gyeonggi-do 15425
REPUBLIC OF KOREA

Name of Object: **Pressure Transmitter
(Digital Pressure Gauge)**

Model(s): **SMAJG, SMAKG, SMANG, SMAOG**

Description of Object: Rated input voltage: 12-24 V d.c.
Rated input current: 60 mA

Tested according to: EN 55011:2016/A11:2020
EN 61326-1:2013
EN IEC 61326-1:2021

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: KR23-YEC0093

Date, 2023-07-27

(Laurent Yuan)

Page 1 of 1

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.



TUV®